



To complete this activity, the participant will need to access and read AORN (2020). Energy Generating Devices. *Guidelines for Perioperative Practice*. Denver, CO: AORN.

Across

- 5. This person has the authority and responsibility for supervising laser safety. (3 words)
- 7. The government agency that mandates the posting of laser warning signs.
- 8. Lasers should be placed in this mode when not in active use.
- 9. The agency that approves corneal eye shields for patients undergoing laser surgery around or near the eye.
- 11. Laser audits must be conducted _____.
- 12. The zone that dictates the use of eye protection.

Down

- 1. The type of laser wavelength that does not have the potential to pass through windows. (2 words)
- 2. Laser eyewear must be labeled with the appropriate optical density and _____.
- 3. The ET tube cuff should be filled with either saline tinted with methylene blue or _____ to decrease the potential for ET tube ignition
- 4. Use _____ lubricant at the surgical site when using lasers. (2 words)
- 6. Sponges should be _____ when used during perineal surgery to prevent methane gas leakage.
- 10. Endotracheal tubes should be laser _____ if the airway is involved.
- 13. This gas must be stopped or decreased when an ignition source is used near the head, neck, or face.

* Content reviewed by the CCI Nursing Education Department for alignment with clinical practice standards. CCI does not require, recommend, or endorse specific training programs in specialized practice areas for any of its exams. This is an example of a future points activity for recertification in collaboration with CCI.